



COLLEGE of AMERICAN  
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

ORIGINAL EVALUATION

TM-A 2023 Tumor Markers

INSTITUTION:

ATTENTION:

CAP NUMBER:

KIT INFORMATION:

COPIED TO:

LEGEND:

Klinikum der Universität München

Munich GE 80336

Tim Kuehnle PhD

8481165-01

Kit# 1

Kit ID:

Kit Mailed:

36136891

1/3/2023

Original Evaluation:

1/31/2023

Next Mailing Date:

7/17/2023

CAP

Exception Reason Codes appearing in this evaluation:  
[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Reviewed By TIT KUEHNLE

Date 02 FEB 2023



**COLLEGE of AMERICAN  
PATHOLOGISTS**  
325 Westlake Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

CAP Number: 8481165-01      Kit# 1  
Institution: Klinikum der Universität München  
Attention: Tim Kuehnle PhD  
City / State: Munich GE 80336

Kit ID: 36136891  
Kit Mailed: 1/3/2023  
Original Evaluation: 1/31/2023

## EVALUATION ORIGINAL

### TM-A 2023 Tumor Markers

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics								Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey      -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower      Upper	Your Grade	
ACTH pg/mL DIASORIN	TM-01	19.3	20.48	1.87	11	-0.6	14.8      26.1	Acceptable	<div>TM-A 2023 TM-B 2022 TM-A 2022</div> <p>-100 -80 -60 -40 -20 0 20 40 60 80 100</p>
	TM-02	194.2	202.01	12.56	11	-0.6	164.3      239.7	Acceptable	
	TM-03	53.0	56.03	3.18	11	-1.0	46.5      65.6	Acceptable	
Thyroglobulin ng/mL OTHER	TM-01	2.0			17			See Note [20]	
	TM-02	38.7			17			See Note [20]	
	TM-03	26.0			17			See Note [20]	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.