



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

VITD-B 2019 25-OH Vitamin D

INSTITUTION:

Klinikum der Universität München
Munich GE 80336
Sarina Benedix

ATTENTION:

CAP NUMBER:

8481165-01

Kit# 1

KIT INFORMATION:

Kit ID:

Kit Mailed:

Original Evaluation:

32632509

11/25/2019

12/24/2019

COPIED TO:

CAP

Next Mailing Date:
4/27/2020

LEGEND:

Exception Reason Codes appearing in this evaluation:

<NONE>

Reviewed By

Date



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01

Kit# 1

Institution: Klinikum der Universität München

Attention: Sarina Benedix

City / State: Munich GE 80336

Kit ID: 32632509

Kit Mailed: 11/25/2019

Original Evaluation: 12/24/2019

EVALUATION ORIGINAL

VITD-B 2019 25-OH Vitamin D

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics										Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+ 100	
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability		Your Grade			
25 OH Vitamin D ng/mL IDS i10, iSYS	VITD-04	18.7	20.04	4.69	13	-0.3	15.0	25.1	Acceptable			
	VITD-05	30.6	31.21	3.04	13	-0.2	23.4	39.1	Acceptable			
	VITD-06	39.2	39.32	5.41	13	0.0	29.4	49.2	Acceptable			