



COLLEGE of AMERICAN
PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

Y-B 2020 Ligand-Special

INSTITUTION:

Klinikum der Universität München
Munich GE 80336

ATTENTION:

Tim Kuehnle PhD

CAP NUMBER:

848165-01

Kit# 1

KIT INFORMATION:

Kit ID:

Kit Mailed:

Original Evaluation:

Next Mailing Date:

32784411

9/21/2020

11/3/2020

4/26/2021

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Reviewed By 



Date 

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.

EVALUATION ORIGINAL

Y-B 2020 Ligand-Special



Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics							Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade
Androstenedione ng/dL OTHER	Y-04	78.00			11			See Note [20]
	Y-05	114.00			10			See Note [20]
	Y-06	24.00			10			See Note [20]
DHEA Sulfate ug/dL OTHER	Y-04	4.7			7			See Note [20]
	Y-05	7.4			7			See Note [20]
	Y-06	1.1			7			See Note [20]
Growth Hormone ng/mL (ug/L) IDS i10, iSYS	Y-04	15.2	15.41	0.55	12	-0.4	13.7 17.1	Acceptable
	Y-05	18.5	18.96	0.59	11	-0.8	17.1 20.8	Acceptable
	Y-06	5.9	5.83	0.18	12	+0.4	5.2 6.4	Acceptable
17-Hydroxyprogesterone ng/dL OTHER	Y-04	408.7			28			See Note [20]
	Y-05	588.0			27			See Note [20]
	Y-06	38.6			27			See Note [20]
IGF-1 ng/mL IDS i10, iSYS	Y-04	71.0	66.45	5.59	25	+0.8	49.6 83.3	Acceptable
	Y-05	85.0	77.43	8.37	25	+0.9	52.3 102.6	Acceptable
	Y-06	25.0	23.76	2.09	24	+0.6	17.5 30.1	Acceptable



The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.

EVALUATION ORIGINAL

Y-B 2020 Ligand-Special

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics										Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability		Your Grade		
Testosterone ng/dL OTHER	Y-04	386.6			4				See Note [20]		
	Y-05	534.8			4				See Note [20]		
	Y-06	58.9			4				See Note [20]		
C-Peptide ng/mL DIASORIN LIAISON	ING-04	1.73	1.354	0.169	11	+2.2	0.84	1.87	Acceptable		
	ING-05	6.30	6.102	0.433	11	+0.5	4.80	7.41	Acceptable		
	ING-06	8.75	8.629	0.723	11	+0.2	6.46	10.80	Acceptable		
Insulin uIU/mL DIASORIN	ING-04	20.0	21.47	2.38	10	-0.6	14.3	28.7	Acceptable		
	ING-05	72.7	70.94	6.44	10	+0.3	51.6	90.3	Acceptable		
	ING-06	242.6	219.83	13.76	10	+1.7	178.5	261.2	Acceptable		
Aldosterone ng/dL DIASORIN LIAISON	RAP-04	42.10	44.567	3.920	60	-0.6	32.80	56.33	Acceptable		
	RAP-05	6.70	7.505	0.846	61	-1.0	4.96	10.05	Acceptable		
	RAP-06	2.90	1.555	0.727	29	+1.8	0.00	3.74	Acceptable		
Renin, direct DIASORIN LIAISON uIU/mL	RAP-04	169.9	170.56	19.50	32	0.0	112.0	229.1	Acceptable		
	RAP-05	78.7	88.42	9.54	32	-1.0	59.8	117.1	Acceptable		
	RAP-06	34.7	36.42	4.71	32	-0.4	22.2	50.6	Acceptable		

x: Result is outside the acceptable limits

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.