



COLLEGE of AMERICAN
PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

VITD-A 2021 25-OH Vitamin D

INSTITUTION:

ATTENTION:

CAP NUMBER:

KIT INFORMATION:

COPIED TO:

LEGEND:

Klinikum der Universitat Munchen

Munich GE 80336

Tim Kuehnle PhD

8481165-01

Kit# 1

Kit ID:

Kit Mailed:

Original Evaluation:

Next Mailing Date:

33828068

4/26/2021

5/24/2021

11/22/2021

CAP

Exception Reason Codes appearing in this evaluation:
<NONE>

Reviewed By 

Date 


The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.

CAP Number: 8481165-01 **Kit#** 1
Institution: Klinikum der Universität München
Attention: Tim Kuehnle PhD
City / State: Munich GE 80336

Kit ID: 33828068
Kit Mailed: 4/26/2021
Original Evaluation: 5/24/2021

EVALUATION ORIGINAL

VTTD-A 2021 25-OH Vitamin D

Test Unit of Measure Ref Group	Evaluation and Comparative Method Statistics							Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	
25 OH Vitamin D ng/mL IDS i10, iSYS	VTTD-01				7			VTTD-A 2021 VTTD-B 2020 
	VTTD-02				7			
	VTTD-03				7			
	All Methods							
	VTTD-01	47.9	49.88	7.55	1897	-0.3	37.4 62.4	Acceptable
	VTTD-02	28.6	27.60	4.11	1904	+0.2	20.6 34.5	Acceptable
	VTTD-03	13.6	12.61	2.51	1856	+0.4	7.6 17.7	Acceptable

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.