COLLEGE of AMERICAN PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

LS-A 2022 Ligand Special

INSTITUTION: Klinikum der Universitat Munchen
Munich GE 80336

ATTENTION: Tim Kuehnle PhD

CAP NUMBER: 8481165-01 **Kit#** 1

KIT INFORMATION:

Kit ID:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
35222910	4/25/2022	6/2/2022	9/19/2022

COPIED TO: CAP

LEGEND: **Exception Reason Codes appearing in this evaluation:**
<NONE>

Reviewed By

20 JUL 2022
Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.





COLLEGE of AMERICAN PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit#** 1
Institution: Klinikum der Universitat Munchen
Attention: Tim Kuehnle PhD
City / State: Munich GE 80336

Kit ID: 35222910
Kit Mailed: 4/25/2022
Original Evaluation: 6/2/2022

EVALUATION
ORIGINAL

LS-A 2022 Ligand Special

Test / Unit of Measure / Peer Group	Specimen	Your Result	Mean	S.D.	No. of Labs	Limits of Acceptability S.D.I	Lower	Upper	Your Grade	Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation
C-Peptide	**ING-01**	6.08	5.953	0.416	10	+0.3	4.70	7.20	Acceptable	
ng/mL	**ING-02**	8.63	8.970	0.666	10	-0.5	6.97	10.97	Acceptable	
DIASORIN LIAISON	**ING-03**	1.72	1.763	0.103	10	-0.4	1.45	2.08	Acceptable	
Insulin	**ING-01**	68.6	73.25	3.31	11	-1.4	63.3	83.2	Acceptable	
uIU/mL	**ING-02**	261.8	257.90	11.88	11	+0.3	222.2	293.6	Acceptable	
DIASORIN	**ING-03**	21.1	22.76	0.86	11	-1.9	20.1	25.4	Acceptable	
Aldosterone	**RAP-01**	2.19	1.800	0.764	27	+0.5	0.00	4.10	Acceptable	
ng/dL	**RAP-02**	36.10	34.254	3.685	68	+0.5	23.20	45.31	Acceptable	
DIASORIN LIAISON	**RAP-03**	6.42	5.333	1.079	66	+1.0	2.09	8.58	Acceptable	
Renin, direct	**RAP-01**	45.0	41.01	3.59	29	+1.1	30.2	51.8	Acceptable	
	RAP-02	308.7	318.50	23.62	29	-0.4	247.6	389.4	Acceptable	
DIASORIN LIAISON	**RAP-03**	98.1	99.58	10.64	28	-0.1	67.6	131.6	Acceptable	
uU/mL										

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.