



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

TM-B 2023 Tumor Markers

INSTITUTION:

Klinikum der Universität München
Munich GE 80336

ATTENTION:

Tim Kuehnle PhD

CAP NUMBER:

8481165-01

KIT INFORMATION:

Kit ID:
36136893

COPIED TO:

CAP

IMPORTANT NOTES:

LEGEND:

Exception Reason Codes appearing in this evaluation:

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Next Mailing Date:
1/2/2024

Original Evaluation:
10/23/2023

Kit Mailed:
9/19/2023

Kit# 1

TIM KUEHNLE

Reviewed By

24 OCT 2023

Date



COLLEGE of AMERICAN
PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01

Kit# 1

Kit ID: 36136893

Institution: Klinikum der Universität München

Kit Mailed: 9/19/2023

Attention: Tim Kuehnle PhD

Original Evaluation: 10/23/2023

City / State: Munich GE 80336

EVALUATION ORIGINAL

TM-B 2023 Tumor Markers

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics								Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade	
ACTH pg/mL DIASORIN	TM-04	197.6	211.80	23.46	16	-0.6	141.4 282.2	Acceptable	
	TM-05	47.5	53.19	9.45	17	-0.6	24.8 81.6	Acceptable	
	TM-06	18.9	19.07	5.66	17	0.0	2.0 36.1	Acceptable	
Thyroglobulin ng/mL OTHER	TM-04	40.0			35			See Note [20]	
	TM-05	21.8			36			See Note [20]	
	TM-06	1.8			36			See Note [20]	

