



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

Y-B 2019 Ligand-Special

INSTITUTION:

Klinikum der Universitat Munchen
Munich GE 80336
Sarina Benedix

ATTENTION:

CAP NUMBER:

8481165-01 Kit# 1

KIT INFORMATION:

Kit ID: 32631496
Kit Mailed: 9/23/2019
Original Evaluation: 10/31/2019

COPIED TO:

CAP

Next Mailing Date:

4/27/2020

LEGEND:

Exception Reason Codes appearing in this evaluation:

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional

[Signature]

Reviewed By

02 DEC 2019

Date

*RESULTS RECEIVED ONLINE
ON 04 NOV 2019*



COLLEGE of AMERICAN PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01

Kit# 1

Institution: Klinikum der Universität München

Attention: Sarina Benedix

City / State: Munich GE 80336

Kit ID: 32631496



Kit Mailed: 9/23/2019

Original Evaluation: 10/31/2019

EVALUATION ORIGINAL



Y-B 2019 Ligand-Special

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics						Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper Your Grade
Androstenedione ng/dL OTHER	Y-04	93.00			13		See Note [20]
	Y-05	115.00			13		See Note [20]
	Y-06	44.00			13		See Note [20]
DHEA Sulfate ug/dL OTHER	Y-04	387.0			9		See Note [20]
	Y-05	500.2			9		See Note [20]
	Y-06	106.6			9		See Note [20]
Growth Hormone ng/mL (ug/L) IDS i10, iSYS	Y-04	17.3	17.89	0.64	11	-0.9	15.9 19.9 Acceptable
	Y-05	20.0	20.55	0.69	11	-0.8	18.4 22.7 Acceptable
	Y-06	7.5	7.66	0.28	11	-0.6	6.8 8.5 Acceptable
17-Hydroxyprogesterone ng/dL OTHER	Y-04	606.0			31		See Note [20]
	Y-05	800.0			31		See Note [20]
	Y-06	50.0			31		See Note [20]
IGF-1 ng/mL IDS i10, iSYS	Y-04	65.0	74.25	5.07	25	-1.8	59.0 89.5 Acceptable
	Y-05	75.0	82.34	5.69	25	-1.3	65.2 99.5 Acceptable
	Y-06	33.0	37.46	3.09	25	-1.4	28.2 46.8 Acceptable



**EVALUATION
ORIGINAL**

Y-B 2019 Ligand-Special

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics					Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation	
	Specimen	Your Result	Mean	S.D.	No. of Labs	Limits of Acceptability Upper Lower	Survey -100-----Mean-----+100
Testosterone ng/dL OTHER	Y-04	601.7			5	See Note [20]	
	Y-05	759.7			5	See Note [20]	
	Y-06	119.9			5	See Note [20]	
SHBG nmol/L nmol/L OTHER	Y-04	71.00			4	See Note [20]	
	Y-05	85.10			4	See Note [20]	
	Y-06	16.67			4	See Note [20]	
C-Peptide ng/mL DIASORIN LIAISON	ING-04	1.84			9	See Note [20]	
	ING-05	8.86			9	See Note [20]	
	ING-06	6.47			9	See Note [20]	
Insulin uIU/mL DIASORIN	ING-04	18.9			8	See Note [20]	 Y-A 2019
	ING-05	211.3			8	See Note [20]	
	ING-06	68.9			8	See Note [20]	
Aldosterone ng/dL DIASORIN LIAISON	RAP-04	2.90	1.693	0.753	29	+1.6 0.00	 Y-B 2019 Y-A 2019
	RAP-05	4.10	39.246	3.268	56	-10.8 29.44	
	RAP-06	2.90	2.935	0.787	43	0.0 0.57	



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01

Kit# 1

Kit ID: 32631496

Institution: Klinikum der Universität München

Kit Mailed: 9/23/2019

Attention: Sarina Benedix

Original Evaluation: 10/31/2019

City / State: Munich GE 80336

EVALUATION
ORIGINAL

Y-B 2019 Ligand-Special

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics							Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation	
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade	Survey -100-----Mean-----+100
Renin, direct DIASORIN LIAISON uU/mL	RAP-04	36.9	43.85	4.70	26	-1.5	29.7	58.0	Y-B 2019 Y-A 2019 x: Result is outside the acceptable limits
	RAP-05	378.0	414.55	38.67	26	-0.9	298.5	530.6	
	RAP-06	99.8	111.34	11.50	26	-1.0	76.8	145.9	