



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

LS-B 2022 Ligand Special

INSTITUTION: Klinikum der Universitat Munchen
Munich GE 80336

ATTENTION: Tim Kuehnle PhD

CAP NUMBER: 8481165-01 Kit# 1

KIT INFORMATION:	Kit ID:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
	35222962	9/19/2022	10/21/2022	4/24/2023

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>

O. Bender
SARINA BENDER

Reviewed By

24.10.22

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.







**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit#** 1
Institution: Klinikum der Universität München
Attention: Tim Kuehnle PhD
City / State: Munich GE 80336

Kit ID: 35222962
Kit Mailed: 9/19/2022
Original Evaluation: 10/21/2022

EVALUATION ORIGINAL

IS-B 2022 Ligand Special

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics										Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade		
C-Peptide ng/mL DIASORIN LIAISON	ING-04	1.64	1.729	0.077	12	-1.2	1.49	1.97	Acceptable		
	ING-05	5.87	6.105	0.191	11	-1.2	5.53	6.68	Acceptable		
	ING-06	8.26	8.436	0.404	12	-0.4	7.22	9.65	Acceptable		
	LS-B 2022 LS-A 2022										
Insulin uIU/mL DIASORIN	ING-04	20.7	21.53	1.18	10	-0.7	17.9	25.1	Acceptable		
	ING-05	69.1	71.07	4.58	10	-0.4	57.3	84.9	Acceptable		
	ING-06	201.0	209.28	10.53	10	-0.8	177.6	240.9	Acceptable		
	LS-B 2022 LS-A 2022										
Aldosterone ng/dL DIASORIN LIAISON	RAP-04	4.88	5.073	0.867	71	-0.2	2.47	7.68	Acceptable		
	RAP-05	29.10	32.176	2.479	71	-1.2	24.73	39.62	Acceptable		
	RAP-06	0.97	1.750	0.881	30	-0.9	0.00	4.40	Acceptable		
	LS-B 2022 LS-A 2022										
Renin, direct uIU/mL DIASORIN LIAISON	RAP-04	77.1	76.16	7.72	33	+0.1	53.0	99.4	Acceptable		
	RAP-05	226.4	209.09	21.19	34	+0.8	145.5	272.7	Acceptable		
	RAP-06	47.1	43.03	4.39	33	+0.9	29.8	56.3	Acceptable		
	LS-B 2022 LS-A 2022										

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

Y-B 2022 Sex Hormones

INSTITUTION:

Klinikum der Universität München
Munich GE 80336

ATTENTION:

Tim Kuehnle PhD

CAP NUMBER:

8481165-01

Kit# 1

KIT INFORMATION:

Kit ID:
35221160

Kit Mailed:
9/19/2022

Original Evaluation:
10/19/2022

Next Mailing Date:
4/24/2023

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:
<NONE>

Interlaboratory test not received!

Reviewed By

D. Bueck
Stefan Bueck

Date

24.10.22