



COLLEGE of AMERICAN
PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

LS-A 2023 Ligand Special

INSTITUTION:

Klinikum der Universität München
Munich GE 80336

ATTENTION:

Tim Kuehnle PhD

CAP NUMBER:

8481165-01

Kit# 1

KIT INFORMATION:

Kit ID:

36136792

Kit Mailed:

4/24/2023

Original Evaluation:

5/25/2023

Next Mailing Date:

9/18/2023

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:
<NONE>

Reviewed By

Jack

Date

06 Jan 2023






**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit#** 1
Institution: Klinikum der Universität München
Attention: Tim Kuehnle PhD
City / State: Munich GE 80336

Kit ID: 36136792
Kit Mailed: 4/24/2023
Original Evaluation: 5/25/2023

EVALUATION ORIGINAL

LS-A 2023 Ligand Special

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics										Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation	
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability		Your Grade	Survey	-100-----Mean-----+100	
C-Peptide ng/mL DIASORIN LIAISON	ING-01	1.61	1.807	0.191	12	-1.0	1.23	2.38	Acceptable	LS-A 2023 LS-B 2022 LS-A 2022		
	ING-02	5.67	6.105	0.316	11	-1.4	5.15	7.06	Acceptable			
	ING-03	8.66	8.953	0.744	12	-0.4	6.72	11.19	Acceptable			
Insulin uIU/mL DIASORIN	ING-01	21.6	22.92	2.13	13	-0.6	16.5	29.3	Acceptable	LS-A 2023 LS-B 2022 LS-A 2022		
	ING-02	64.3	73.91	6.44	13	-1.5	54.5	93.3	Acceptable			
	ING-03	243.0	263.34	23.67	13	-0.9	192.3	334.4	Acceptable			
Aldosterone ng/dL DIASORIN LIAISON	RAP-01	4.34	4.903	0.953	76	-0.6	2.04	7.77	Acceptable	LS-A 2023 LS-B 2022 LS-A 2022		
	RAP-02	0.97	1.834	0.929	32	-0.9	0.00	4.63	Acceptable			
	RAP-03	37.20	33.351	3.084	74	+1.2	24.10	42.61	Acceptable			
Renin, direct DIASORIN LIAISON uU/mL	RAP-01	75.9	74.73	7.58	33	+0.2	51.9	97.5	Acceptable	LS-A 2023 LS-B 2022 LS-A 2022		
	RAP-02	45.2	42.06	5.03	33	+0.6	26.9	57.2	Acceptable			
	RAP-03	216.7	205.82	15.97	30	+0.7	157.9	253.8	Acceptable			