



COLLEGE of AMERICAN
PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

VITD-B 2021 25-OH Vitamin D

INSTITUTION:

Klinikum der Universität München
Munich GE 80336

ATTENTION:

Tim Kuehnle PhD

CAP NUMBER:

8481165-01

Kit# 1

KIT INFORMATION:

Kit ID:
33828077

Kit Mailed:
11/22/2021

Original Evaluation:
12/16/2021


Next Mailing Date:
4/25/2022

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:
<NONE>



Reviewed By


Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.

EVALUATION ORIGINAL

VITD-B 2021 25-OH Vitamin D

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics								Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade	
25 OH Vitamin D ng/mL IDS i10, iSYS	VITD-04				6				VITD-B 2021 VITD-A 2021 VITD-B 2020 
	VITD-05				7				
	VITD-06				7				
	All Methods								
	VITD-04	16.8	15.32	3.74	2034	+0.4	10.3 20.4	Acceptable	
	VITD-05	53.2	48.55	6.68	2004	+0.7	36.4 60.7	Acceptable	
	VITD-06	35.7	33.18	4.23	2043	+0.6	24.8 41.5	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.