



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

TM-A 2020 Tumor Markers

INSTITUTION:

Klinikum der Universität München
Munich GE 80336
Sarina Benedix

ATTENTION:

CAP NUMBER:

8481165-01 Kit# 1

KIT INFORMATION:

Kit ID: 32786362
Kit Mailed: 1/6/2020

COPIED TO:

CAP

Next Mailing Date:
7/20/2020

Original Evaluation:
2/18/2020

LEGEND:

Exception Reason Codes appearing in this evaluation:

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Tina Munnich

Reviewed By

06 MAR 2020

Date



COLLEGE of AMERICAN PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 Kit# 1

Institution: Klinikum der Universität München

Attention: Sarina Benedix

City / State: Munich GE 80336

Kit ID: 32786362

Kit Mailed: 1/6/2020

Original Evaluation: 2/18/2020

EVALUATION

ORIGINAL

TM-A 2020 Tumor Markers

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics								Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+ 100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade	
ACTH pg/mL DIASORIN	TM-01	185.8	187.85	17.98	11	-0.1	133.8 241.8	Acceptable	
	TM-02	25.1	28.77	3.79	11	-1.0	17.4 40.2	Acceptable	
	TM-03	17.8	23.25	6.28	11	-0.9	4.4 42.1	Acceptable	
Thyroglobulin ng/mL OTHER	TM-01	76.8			11			See Note [20]	
	TM-02	38.1			12			See Note [20]	
	TM-03	2.6			11			See Note [20]	

