



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4060 • cap.org

ORIGINAL EVALUATION

TM-B 2022 Tumor Markers

INSTITUTION:

Klinikum der Universität München
Munich GE 80336

ATTENTION:

Tim Kuehnle PhD

CAP NUMBER:

8481165-01

Kit# 1

KIT INFORMATION:

Kit ID:

Kit Mailed:

35210310

7/18/2022

Original Evaluation:

8/18/2022

Next Mailing Date:

1/3/2023

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Tim Kuehnle
Reviewed By

22 Aug 2022
Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit#** 1
Institution: Klinikum der Universitat Munchen
Attention: Tim Kuehnle PhD
City / State: Munich GE 80336

Kit ID: 35210310
Kit Mailed: 7/18/2022
Original Evaluation: 8/18/2022

EVALUATION ORIGINAL

TM-B 2022 Tumor Markers

ORIGINAL DATA										Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation
Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Survey
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability		Your Grade	
ACTH pg/mL, DIASORIN	TM-04	10.5	11.22	1.28	16	-0.6	7.3	15.1	Acceptable	TM-B 2022 TM-A 2022 TM-B 2021
	TM-05	5.6	5.98	1.19	15	-0.3	2.4	9.6	Acceptable	
	TM-06	122.8	136.89	11.64	15	-1.2	101.9	171.9	Acceptable	
Thyroglobulin ng/mL, OTHER	TM-04	20.0			26				See Note [20]	<div>-100 -80 -60 -40 -20 0 20 40 60 80 100</div>
	TM-05	1.7			24				See Note [20]	
	TM-06	41.1			24				See Note [20]	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.