



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

Y-B 2023 Sex Hormones

INSTITUTION:

Klinikum der Universität München
Munich GE 80336
Tim Kuehnle PhD

ATTENTION:

CAP NUMBER:

8481165-01 Kit# 1

KIT INFORMATION:

Kit ID: 36136862
Kit Mailed: 9/18/2023

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Next Mailing Date:
4/22/2024

Original Evaluation:
11/1/2023

[Signature]

Tim Kuehnle

Reviewed By

Date

02 NOV 2023

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



COLLEGE of AMERICAN
PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60063-2750
800-323-4040 • cap.org

CAP Number: 8481165-01

Kit# 1

Kit ID: 36136862

Institution: Klinikum der Universität München

Kit Mailed: 9/18/2023


Attention: Tim Kuehnle PhD

Original Evaluation: 11/1/2023

City / State: Munich GE 80336

EVALUATION ORIGINAL

Y-B 2023 Sex Hormones

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics							Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation	
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade	Survey -100-----Mean-----+100
Androstenedione ng/dL OTHER	Y-04	240.00			10			See Note [20]	
	Y-05	285.00			12			See Note [20]	
	Y-06	57.70			12			See Note [20]	
DHEA Sulfate ug/dL OTHER	Y-04	109.7			5			See Note [20]	
	Y-05	413.7			5			See Note [20]	
	Y-06	373.7			5			See Note [20]	
Growth Hormone ng/mL (ug/L) IDS i10, iSYS	Y-04	9.5	9.05	0.35	14	+1.3	8.0	10.1 Acceptable	
	Y-05	20.1	19.38	0.60	14	+1.2	17.5	21.2 Acceptable	
	Y-06	17.7	17.01	0.62	14	+1.1	15.1	18.9 Acceptable	
17-Hydroxyprogesterone ng/dL IDS i10, iSYS	Y-04	92.0			7			See Note [20]	
	Y-05	284.0			7			See Note [20]	
	Y-06	400.0			7			See Note [20]	
IGF-1 ng/mL IDS i10, iSYS	Y-04	48.0	47.57	2.55	20	+0.2	39.9	55.3 Acceptable	
	Y-05	69.0	70.34	3.57	20	-0.4	59.6	81.1 Acceptable	
	Y-06	85.0	87.37	4.48	20	-0.5	73.9	100.9 Acceptable	



COLLEGE of AMERICAN
PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 Kit# 1
Institution: Klinikum der Universität München
Attention: Tim Kuehnle PhD
City / State: Munich GE 80336

Kit ID: 36136862
Kit Mailed: 9/18/2023
Original Evaluation: 11/1/2023

EVALUATION
ORIGINAL

Y-B 2023 Sex Hormones

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics							Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper Your Grade	
Testosterone ng/dL OTHER	Y-04	116.9			4		See Note [20]	
	Y-05	948.7			4		See Note [20]	
	Y-06	546.9			4		See Note [20]	
SHBG nmol/L nmol/L OTHER	Y-04	35.76			5		See Note [20]	
	Y-05	37.40			5		See Note [20]	
	Y-06	86.56			5		See Note [20]	