



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

VITD-A 2019 25-OH Vitamin D

INSTITUTION:

Klinikum der Universitat Munchen
Munich GE 80336
Sarina Benedix

ATTENTION:

CAP NUMBER:

8481165-01 Kit# 1

KIT INFORMATION:

Kit ID: 32632504
Kit Mailed: 4/29/2019
Original Evaluation: 5/29/2019

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:
<NONE>

Next Mailing Date:
11/25/2019

J. Clin

Reviewed By

31 MAY 2019

SF 31 MAY 2019

28 AUG 2019

Date



COLLEGE of AMERICAN
PATHOLOGISTS
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 Kit# 1
Institution: Klinikum der Universität München
Attention: Sarina Benedix
City / State: Munich GE 80336

Kit ID: 32632504
Kit Mailed: 4/29/2019
Original Evaluation: 5/29/2019

EVALUATION ORIGINAL

VITD-A 2019 25-OH Vitamin D

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics						Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation	
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade
25 OH Vitamin D ng/mL IDS iSYS	VITD-01	15.6	14.46	1.83	16	+0.6	9.4 19.5	Acceptable
	VITD-02	45.8	48.51	4.36	16	-0.6	36.3 60.7	Acceptable
	VITD-03	30.2	31.22	3.22	16	-0.3	23.4 39.1	Acceptable
Survey -100-----Mean-----+100 VITD-A 2019 -100 -80 -60 -40 -20 0 20 40 60 80 100								