



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

TM-B 2019 Tumor Markers

INSTITUTION:

Klinikum der Universität München
Munich GE 80336
Sarina Benedix

ATTENTION:

CAP NUMBER:

8481165-01 Kit# 1

KIT INFORMATION:

Kit ID: 32632533
Kit Mailed: 7/22/2019

COPIED TO:

CAP

Next Mailing Date:
1/6/2020

Original Evaluation:
8/27/2019

LEGEND:

Exception Reason Codes appearing in this evaluation:

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

[Signature]

Reviewed By

19 SEP 2019

Date



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit#** 1
Institution: Klinikum der Universität München
Attention: Sarina Benedix
City / State: Munich GE 80336

Kit ID: 32632533
Kit Mailed: 7/22/2019
Original Evaluation: 8/27/2019

EVALUATION ORIGINAL

TM-B 2019 Tumor Markers

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics							Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper	Your Grade
ACTH pg/mL DIASORIN	TM-04	33.3	27.31	5.38	13	+1.1	11.1	43.5 Acceptable
	TM-05	223.5	199.19	10.12	14	+2.4	168.8	229.6 Acceptable
	TM-06	42.0	37.26	4.38	14	+1.1	24.1	50.4 Acceptable
Thyroglobulin ng/mL OTHER	TM-04	0.4			8			See Note [20]
	TM-05	63.9			12			See Note [20]
	TM-06	31.1			12			See Note [20]

